

TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD
ZERTIFIKAT ◆ CERTIFICATE ◆ CERTIFICADO ◆ CERTIFIKAT ◆ 認證書 ◆ CERTIFICATE ◆ CERTIFICATE



CERTIFICATE

No. U8 063179 0074 Rev. 00

Holder of Certificate: **STORZ & BICKEL GMBH & CO. KG**
In Grubenäcker 5-9
78532 Tuttlingen
GERMANY

Production Facility(ies): 063179
Certification Mark:



Product: **General Medical Devices**
Vaporizer system

Brand Name: **vapormed**
Model(s): **VOLCANO MEDIC 2**

Parameters:
Rated Input Voltage: 110- 120 V 220-240 V
Rated Frequency: 50-60 Hz
Rated Input Power: 270 VA
Protection Class: II
Applied Part Type: BF
Protection against Ingress of Water: IP 21

Remarks:
The certificate is valid for USA and Canada.
Other countries might have other requirements,
which were not part of this certification.
ANSI/AAMI ES60601-1:2005/(R)2012
CAN/CSA-C22.2 No. 60601-1:2014

Tested according to:

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited Certification body.

Test report no.: 713115873

Date, 2019-08-26

(Paul Bielawski)

